

AU400 Contamination Prevention Parameters							
No.	PRECEDING TEST NAME	FOLLOWING TEST NAME	REAGENT PROBE CLEANER	WASH COUNT	CANCEL	PREVENT USE	
						Mixer	Cuvette
1.	HCY B08176 R1	Lip 030 R1	CLN 100%*	5	No	Yes	No
2.	HCY B08176 R2	Lip 030 R1	CLN 100%*	5	No	Yes	No
3.	HCY B08176 R1	Lip 030 R2	CLN 100%*	5	No	Yes	No
4.	HCY B08176 R2	Lip 030 R2	CLN 100%*	5	No	Yes	No
5.	UIBC 205 R1	HCY B08176 R1	Water	1	Yes	Yes	No
6.	UIBC 205 R2	HCY B08176 R2	Water	1	Yes	Yes	No
7.	HCY B08176 R1	UIBC 205 R1	Water	1	Yes	Yes	No
8.	HCY B08176 R2	UIBC 205 R2	Water	1	Yes	Yes	No
9.	HCY B08176 R1	LDH 026 R2	Water	1	Yes	Yes	Yes
10.	HCY B08176 R1	LDH 027 R1	Water	1	Yes	Yes	Yes
11.	HCY B08176 R2	LDH 027 R2	Water	1	Yes	Yes	Yes
12.	HCY B08176 R1	LDH 028 R1	Water	1	Yes	Yes	Yes
13.	HCY B08176 R2	LDH 028 R2	Water	1	Yes	Yes	Yes

\* CLN OSR0001/ODR2000