

AU480 Contamination Prevention Parameters							
No.	PRECEDING TEST NAME	FOLLOWING TEST NAME	REAGENT PROBE CLEANER	WASH COUNT	EFFECTIVE OF WATER CLEANING	PREVENT USE	
						Mixer	Cuvette
1.	HCY B08176 R1	Lip 030 R1	CLN 100%*	5	No	Yes	No
2.	HCY B08176 R2	Lip 030 R1	CLN 100%*	5	No	Yes	No
3.	HCY B08176 R1	Lip 030 R2	CLN 100%*	5	No	Yes	No
4.	UIBC 205 R1	HCY B08176 R1	Water	1	Yes	Yes	No
5.	HCY B08176 R1	UIBC 205 R1	Water	1	Yes	Yes	No
6.	HCY B08176 R1	LDH 026 R2	Water	1	Yes	Yes	Yes
7.	HCY B08176 R1	LDH 027 R1	Water	1	Yes	Yes	Yes
8.	HCY B08176 R1	LDH 028 R1	Water	1	Yes	Yes	Yes

\* CLN OSR0001/ODR2000, replenish cleaner on a daily basis or as requested.

Note: When selecting a preceding R1 and a following R1, R2 to R2 is also automatically selected. There is no facility for selecting a preceding R2 and a following R2 on the AU480 instrument.