

AU5800 Contamination Prevention Parameters							
No.	PRECEDING TEST NAME	FOLLOWING TEST NAME	REAGENT PROBE CLEANER	WASH COUNT	EFFECTIVE OF WATER CLEANING	PREVENT USE	
						Mixer	Cuvette
1.	HCY B08176	LIP 030 [*]	N/A				
2.	UIBC 205	HCY B08176	Water	1	Yes	Yes	No
3.	HCY B08176	UIBC 205	Water	1	Yes	Yes	No
4.	HCY B08176	LDH 026	Water	1	Yes	Yes	Yes
5.	HCY B08176	LDH 027	Water	1	Yes	Yes	Yes
6.	HCY B08176	LDH 028	Water	1	Yes	Yes	Yes

^{*}separate Lipase by using designated rings.